



**PRECISION
MEASUREMENTS
INC.**

CERTIFICATE OF CALIBRATION

Report No. 1895977

CUSTOMER

Old ABC, Company
333 Moffett Park Drive
Sunnyvale, CA
94089

Cal Date: 11/18/2011
Due Date: 11/18/2012
Cal Int: 12 Mo.

INSTRUMENT/ID

Manufacturer: Fluke

Serial No: 916704242

Acct No: 4532

Model No: 179

Asset/ID No: 4567

PMC No: 2010

Description: Multimeter, Digital

Department: Engineering

Code No: 00

Location: Bldg. 1

CALIBRATION CONDITIONS

Received: In Tolerance

Cal Spec: Manufacturer

Temp: 22 °C

Returned: In Tolerance

Cal Procedure: 179.pdf

Humidity: 48 %

Cal'd At: PMI Lab

CALIBRATION EQUIPMENT USED

Asset No.	Manufacturer	Model No.	Calib. Date	Due Date
2752	Fluke	5520A	09/09/11	09/09/12

REMARKS/COMMENTS

PRECISION MEASUREMENTS, INC. Quality System complies with ISO 9001:2008 and certifies that all calibration has been performed using standards traceable to the NATIONAL INSTITUTE of STANDARDS and TECHNOLOGY (NIST), or an acceptable value of a natural physical constant or ratio Calibration technique per MIL-STD-45662A, or current versions of ANSI-Z-540 , or ISO/IEC 17025 as applicable. Uncertainty of Standards has a 4:1 ratio or greater than the instrument under test or otherwise accounted for by analysis or quantitatively documented on this Certificate. This certificate applies only to the instrument identified above and shall not be reproduced, except in full, without specific written approval by PRECISION MEASUREMENTS, INC.

TECHNICIAN: Miller By Miller

QC: Shimeg

PM-COC-102A

•333 Moffett Park Drive • Sunnyvale, CA 94089 • (408) 733-8600•